



DIGI

SERVICE MANUAL

MODEL:

DC-120

EDITION

	MONTH	YEAR
1ST	AUGUST	1989
2ND		
3RD		
4TH		
5TH		
6TH		
7TH		
8TH		
9TH		
10TH		

**DIGI MATEX, INC.**

84 SOUTH FRONT STREET  
BERGENFIELD, NJ 07621

201-387-7755

FAX: 201-387-9385

Message List

MESSAGE	CONTENTS
---Add XX	Sample size is too small
All CLEAR	Clearing all items in memory
C XX	Number of items in memory
CH XX	Checking part code
CLEAR	Clearing one item in memory
dA - XX - XX - XX	Setting date and time
dP - 80 Conn on	DP-80 Printer on
dP - 80 Conn off	DP-80 Printer off
Err -	Error during data transfer
InSUFS	Setting insufficient sample size
InVENT CLEAR	Clearing inventory data
Lo-Err	Zero is out of range on the low side
L-YEAR	Setting Leap Year constant
P-no CLEAR	Clearing part number
ProG.	In programming mode
	Transferring data from/to cassette tape
SEt 1	Setting setpoint 1
SEt 2	Setting setpoint 2
SPEC XX	Setting Spec Data
tArE CLEAR	Clearing tare weight in memory
total XXXX	Doing "+" "-" accumulating operation
t - C XX XX	Programming alpha/numeric part number
UnIt CLEAR	Clearing unit weight in memory
UP-Err	Zero is out of range on the high side
UEr - XXX	Software version number

(X: Numerics)

This is a “Table of Contents preview” for quality assurance

The full manual can be purchased from our store:

[https://the-checkout-tech.com/manuals/Digi/DC-120 service and calibration manual.html](https://the-checkout-tech.com/manuals/Digi/DC-120%20service%20and%20calibration%20manual.html)

And our free Online Keysheet maker:

<https://the-checkout-tech.com/Cash-Register-Keysheet-Creator/>

[HTTPS://THE-CHECKOUT-TECH.COM](https://THE-CHECKOUT-TECH.COM)